

7601.2010 SENSITIVITY REQUIREMENT.

For a scale not equipped with a balance indicator, but with a minimum graduated interval less than ten pounds, the sensitivity requirement is three times the value of the minimum graduated interval or 15 pounds, whichever is less.

Statutory Authority: *MS s 239.06*

History: *20 SR 1928*

Published Electronically: *October 2, 2007*